

SUPERVISOR TEST




WARREN COUNTY
 SHERRIF'S OFFICE
 DATAMASTER dmt: 144506

Date: 11/27/2017

OPERATOR NAME:
 TYLER J MORSE
 Permit #: 41959
 Operator Agency/Dept: WARREN CO. SO
 Reference Standard #: 17220

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:02
INTERNAL STANDARD	VERIFIED	09:02
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	09:02
BLANK TEST	0.00	09:03
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	09:04
BLANK TEST	0.00	09:04
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:05
BLANK TEST	0.00	09:06



POJL #440