

SOLUTION CHANGE



 WARREN COUNTY
 SHERRIF'S OFFICE
 DATAMASTER dmt: 144506

Date: 12/01/2017

OPERATOR NAME:
 CHARLES B GEISLER
 Permit #: 33757
 Operator Agency/Dept: WARREN CO. SO
 Reference Standard #: 17310

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:32
INTERNAL STANDARD	VERIFIED	22:33
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	22:33
BLANK TEST	0.00	22:34
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	22:34
BLANK TEST	0.00	22:35
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	22:36
BLANK TEST	0.00	22:36



Charles B. Geisler #241