

SOLUTION CHANGE

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 01/04/2018

OPERATOR NAME:
CHARLES B GEISLER
Permit #: 33757
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 17310

Nominal = 0.10 g/210L

BLANK TEST	0.00	05:39
INTERNAL STANDARD	VERIFIED	05:39
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	05:40
BLANK TEST	0.00	05:41
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	05:41
BLANK TEST	0.00	05:42
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	05:43
BLANK TEST	0.00	05:43

Charles B. Geisler #241

