

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 01/19/2018

OPERATOR NAME:
CHARLES B GEISLER
Permit #: 33757
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 17310

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:09
INTERNAL STANDARD	VERIFIED	01:09
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	01:10
BLANK TEST	0.00	01:11
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	01:11
BLANK TEST	0.00	01:12
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	01:12
BLANK TEST	0.00	01:13

Charles B. Geisler #2416

