

SUPERVISOR TEST

WARREN CO.
 SHERIFF'S OFFICE
 DATAMASTER dmt: 105208

Date: 02/24/2018

OPERATOR NAME:
 CHARLES B GEISLER
 Permit #: 33757
 Operator Agency/Dept: WARREN CO. SO
 Reference Standard #: 17390

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:31
INTERNAL STANDARD	VERIFIED	00:31
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	00:31
BLANK TEST	0.00	00:32
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	00:33
BLANK TEST	0.00	00:34
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	00:34
BLANK TEST	0.00	00:35

Charles B Geisler #241

