

SOLUTION CHANGE

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 05/02/2018

OPERATOR NAME:
RYAN C SCHROECK
Permit #: 41316
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 18040

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:12
INTERNAL STANDARD	VERIFIED	10:12
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	10:13
BLANK TEST	0.00	10:14
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	10:14
BLANK TEST	0.00	10:15
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	10:15
BLANK TEST	0.00	10:16

RCS

