

SOLUTION CHANGE

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 07/01/2018

OPERATOR NAME:
CHARLES B GEISLER
Permit #: 33757
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 18100

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:23
INTERNAL STANDARD	VERIFIED	03:23
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	03:24
BLANK TEST	0.00	03:25
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	03:25
BLANK TEST	0.00	03:26
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	03:26
BLANK TEST	0.00	03:27

Charles B. Geisler

