

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 07/08/2018

OPERATOR NAME:
J T COON
Permit #: 36043
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 18100

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:28
INTERNAL STANDARD	VERIFIED	20:28
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	20:28
BLANK TEST	0.00	20:29
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	20:30
BLANK TEST	0.00	20:31
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	20:31
BLANK TEST	0.00	20:32

J T Coon
341

