

SOLUTION CHANGE

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 07/23/2018

OPERATOR NAME:
RYAN C SCHROECK
Permit #: 41316
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 18100

Nominal = 0.10 g/210L

BLANK TEST	0.00	16:36
INTERNAL STANDARD	VERIFIED	16:36
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	16:36
BLANK TEST	0.00	16:37
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	16:38
BLANK TEST	0.00	16:39
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	16:39
BLANK TEST	0.00	16:40

RC Schroeck

