

SUPERVISOR TEST

WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 105208

Date: 08/22/2018

OPERATOR NAME:  
J T COON  
Permit #: 36043  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 18170

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:00
INTERNAL STANDARD	VERIFIED	23:00
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	23:00
BLANK TEST	0.00	23:01
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	23:02
BLANK TEST	0.00	23:03
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	23:03
BLANK TEST	0.00	23:04

*P. Coon* 341

