

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 08/30/2018

OPERATOR NAME:
CHARLES B GEISLER
Permit #: 33757
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 18170

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:23
INTERNAL STANDARD	VERIFIED	04:23
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	04:24
BLANK TEST	0.00	04:25
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	04:25
BLANK TEST	0.00	04:26
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	04:26
BLANK TEST	0.00	04:27

