

SOLUTION CHANGE

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 09/04/2018

OPERATOR NAME:
CHARLES B GEISLER
Permit #: 33757
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 18170

Nominal = 0.10 g/210L

BLANK TEST	0.00	05:31
INTERNAL STANDARD	VERIFIED	05:31
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	05:31
BLANK TEST	0.00	05:32
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	05:32
BLANK TEST	0.00	05:33
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	05:34
BLANK TEST	0.00	05:35

Charles B. Geisler #244

