

SUPERVISOR TEST

WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 09/22/2018

OPERATOR NAME:
CHARLES B GEISLER
Permit #: 33757
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 18170

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:58
INTERNAL STANDARD	VERIFIED	02:58
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	02:58
BLANK TEST	0.00	02:59
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	02:59
BLANK TEST	0.00	03:00
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	03:01
BLANK TEST	0.00	03:02

Charles B. Geisler #241

