

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 10/15/2018

OPERATOR NAME:
J T COON
Permit #: 36043
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 18230

Nominal = 0.10 g/210L

BLANK TEST	0.00	05:10
INTERNAL STANDARD	VERIFIED	05:10
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	05:11
BLANK TEST	0.00	05:12
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	05:12
BLANK TEST	0.00	05:13
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	05:13
BLANK TEST	0.00	05:14

John T. Coon 341

