

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106606

Date: 03/15/2018

OPERATOR NAME:
CHARLES B GEISLER
Permit #: 33757
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 17390

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:32
INTERNAL STANDARD	VERIFIED	03:32
REFERENCE STANDARD TEMP	33.95c	
REFERENCE STANDARD	0.09	03:32
BLANK TEST	0.00	03:33
REFERENCE STANDARD TEMP	33.95c	
REFERENCE STANDARD	0.09	03:33
BLANK TEST	0.00	03:34
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	03:35
BLANK TEST	0.00	03:36

Handwritten signature: C. B. Geisler #141

