

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106606

Date: 04/28/2018

OPERATOR NAME:
J T COON
Permit #: 36043
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 18040

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:20
INTERNAL STANDARD	VERIFIED	22:20
REFERENCE STANDARD TEMP	34.07c	
REFERENCE STANDARD	0.09	22:21
BLANK TEST	0.00	22:21
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	22:22
BLANK TEST	0.00	22:23
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	22:23
BLANK TEST	0.00	22:24

