

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106606

Date: 05/22/2018

OPERATOR NAME:
J T COON
Permit #: 36043
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 18040

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:44
INTERNAL STANDARD	VERIFIED	22:44
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	22:45
BLANK TEST	0.00	22:46
REFERENCE STANDARD TEMP	33.96c	
REFERENCE STANDARD	0.09	22:46
BLANK TEST	0.00	22:47
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	22:47
BLANK TEST	0.00	22:48



Page 341