

SUPERVISOR TEST

WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106606

Date: 05/29/2018

OPERATOR NAME:  
CHARLES B GEISLER  
Permit #: 33757  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 18040

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:11
INTERNAL STANDARD	VERIFIED	00:11
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	00:11
BLANK TEST	0.00	00:12
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	00:12
BLANK TEST	0.00	00:13
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	00:13
BLANK TEST	0.00	00:14

