

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106606

Date: 07/08/2018

OPERATOR NAME:
J T COON
Permit #: 36043
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 18100

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:03
INTERNAL STANDARD	VERIFIED	21:03
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	21:04
BLANK TEST	0.00	21:05
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	21:05
BLANK TEST	0.00	21:06
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	21:06
BLANK TEST	0.00	21:07

J T Coon

