

SUPERVISOR TEST

WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106606

Date: 08/08/2018

OPERATOR NAME:  
CHARLES B GEISLER  
Permit #: 33757  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 18170

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:31
INTERNAL STANDARD	VERIFIED	01:32
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	01:32
BLANK TEST	0.00	01:33
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	01:33
BLANK TEST	0.00	01:34
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	01:34
BLANK TEST	0.00	01:35

*Charles B Geisler* #241

