

SUPERVISOR TEST

WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106606

Date: 08/23/2018

OPERATOR NAME:  
J T COON  
Permit #: 36043  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 18170

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:56
INTERNAL STANDARD	VERIFIED	00:56
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	00:56
BLANK TEST	0.00	00:57
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	00:57
BLANK TEST	0.00	00:58
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	00:59
BLANK TEST	0.00	01:00

*P.O. [Signature]* 341

