

SOLUTION CHANGE




WARREN COUNTY
SHERRIF'S OFFICE
DATAMASTER dmt: 144506

Date: 01/04/2018

OPERATOR NAME:
CHARLES B GEISLER
Permit #: 33757
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 17310

Nominal = 0.10 g/210L

BLANK TEST	0.00	05:00
INTERNAL STANDARD	VERIFIED	05:00
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	05:01
BLANK TEST	0.00	05:02
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	05:02
BLANK TEST	0.00	05:03
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	05:03
BLANK TEST	0.00	05:04



Charles B. Geisler #224