

SOLUTION CHANGE




WARREN COUNTY
SHERRIF'S OFFICE
DATAMASTER dmt: 144506

Date: 02/01/2018

OPERATOR NAME:
TYLER J MORSE
Permit #: 41959
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 17390

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:49
INTERNAL STANDARD	VERIFIED	10:49
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	10:49
BLANK TEST	0.00	10:50
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	10:50
BLANK TEST	0.00	10:51
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	10:52
BLANK TEST	0.00	10:53



2018 #747

SOLUTION CHANGE

WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506


Date: 02/01/2018

OPERATOR NAME:
TYLER J MORSE
Permit #: 41959
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 17390

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:44
INTERNAL STANDARD	VERIFIED	10:44
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	INTERFERENCE	10:44

Interference Detected



SOLUTION CHANGE

WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 02/01/2018

OPERATOR NAME:
TYLER J MORSE
Permit #: 41959
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 17390

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:32
INTERNAL STANDARD	VERIFIED	10:32
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	INTERFERENCE	10:33

Interference Detected

