

SOLUTION CHANGE



WARREN COUNTY  
SHERRIF'S OFFICE  
DATAMASTER dmt: 144506

Date: 03/30/2018

OPERATOR NAME:  
CHARLES B GEISLER  
Permit #: 33757  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 18040

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:42
INTERNAL STANDARD	VERIFIED	04:42
REFERENCE STANDARD TEMP	33.96c	
REFERENCE STANDARD	0.09	04:42
BLANK TEST	0.00	04:43
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	04:43
BLANK TEST	0.00	04:44
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	04:45
BLANK TEST	0.00	04:45



*Charles B. Geisler*