

SUPERVISOR TEST



WARREN COUNTY
 SHERRIF'S OFFICE
 DATAMASTER dmt: 144506

Date: 04/21/2018

OPERATOR NAME:
 CHARLES B GEISLER
 Permit #: 33757
 Operator Agency/Dept: WARREN CO. SO
 Reference Standard #: 18040

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:15
INTERNAL STANDARD	VERIFIED	23:15
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	23:16
BLANK TEST	0.00	23:16
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	23:17
BLANK TEST	0.00	23:18
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	23:18
BLANK TEST	0.00	23:19



Charles B Geisler #241