

SUPERVISOR TEST




WARREN COUNTY
SHERRIF'S OFFICE
DATAMASTER dmt: 144506

Date: 04/28/2018

OPERATOR NAME:
J T COON
Permit #: 36043
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 18040

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:55
INTERNAL STANDARD	VERIFIED	19:55
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	19:55
BLANK TEST	0.00	19:56
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	19:56
BLANK TEST	0.00	19:57
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	19:57
BLANK TEST	0.00	19:58



P. O. Coon 341