

SOLUTION CHANGE

 WARREN COUNTY
 SHERRIF'S OFFICE
 DATAMASTER dmt: 144506

Date: 05/02/2018

OPERATOR NAME:
 RYAN C SCHROECK
 Permit #: 41316
 Operator Agency/Dept: WARREN CO. SO
 Reference Standard #: 18040

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:51
INTERNAL STANDARD	VERIFIED	10:51
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	10:51
BLANK TEST	0.00	10:52
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	10:52
BLANK TEST	0.00	10:53
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	10:54
BLANK TEST	0.00	10:55

90 R. Schrock

