

SUPERVISOR TEST




WARREN COUNTY  
SHERRIF'S OFFICE  
DATAMASTER dmt: 144506

Date: 05/08/2018

OPERATOR NAME:  
J T COON  
Permit #: 36043  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 18040

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:03
INTERNAL STANDARD	VERIFIED	23:03
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	23:03
BLANK TEST	0.00	23:04
REFERENCE STANDARD TEMP	33.96c	
REFERENCE STANDARD	0.09	23:04
BLANK TEST	0.00	23:05
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	23:06
BLANK TEST	0.00	23:07



*P. Coon*  
341