

SUPERVISOR TEST




WARREN COUNTY
SHERRIF'S OFFICE
DATAMASTER dmt: 144506

Date: 05/15/2018

OPERATOR NAME:
TYLER J MORSE
Permit #: 41959
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 18040

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:05
INTERNAL STANDARD	VERIFIED	13:05
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	13:06
BLANK TEST	0.00	13:07
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	13:07
BLANK TEST	0.00	13:08
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	13:08
BLANK TEST	0.00	13:09



POH #14412