

SUPERVISOR TEST

WARREN COUNTY
SHERRIF'S OFFICE
DATAMASTER dmt: 144506

Date: 06/21/2018

OPERATOR NAME:
CHARLES B GEISLER
Permit #: 33757
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 18100

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:06
INTERNAL STANDARD	VERIFIED	23:06
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	23:07
BLANK TEST	0.00	23:08
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	23:08
BLANK TEST	0.00	23:09
REFERENCE STANDARD TEMP	33.96c	
REFERENCE STANDARD	0.09	23:09
BLANK TEST	0.00	23:10

