

SOLUTION CHANGE



WARREN COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 144506

Date: 07/01/2018

OPERATOR NAME:  
CHARLES B GEISLER  
Permit #: 33757  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 18100

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:49
INTERNAL STANDARD	VERIFIED	02:49
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	02:49
BLANK TEST	0.00	02:50
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	02:51
BLANK TEST	0.00	02:52
REFERENCE STANDARD TEMP	33.95c	
REFERENCE STANDARD	0.09	02:52
BLANK TEST	0.00	02:53



*Sgt. Charles B. Geisler*