

SUPERVISOR TEST

WARREN COUNTY  
SHERRIF'S OFFICE  
DATAMASTER dmt: 144506

Date: 07/08/2018

OPERATOR NAME:  
J T COON  
Permit #: 36043  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 18100

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:10
INTERNAL STANDARD	VERIFIED	20:10
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	20:10
BLANK TEST	0.00	20:11
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	20:12
BLANK TEST	0.00	20:13
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	20:13
BLANK TEST	0.00	20:14

*J T Coon*  
301

