

SOLUTION CHANGE




WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 08/03/2018

OPERATOR NAME:
CHARLES B GEISLER
Permit #: 33757
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 18170

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:48
INTERNAL STANDARD	VERIFIED	03:48
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	03:49
BLANK TEST	0.00	03:49
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	03:50
BLANK TEST	0.00	03:51
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	03:51
BLANK TEST	0.00	03:52



Charles B. Geisler #241