

SOLUTION CHANGE

WARREN COUNTY
 SHERRIF'S OFFICE
 DATAMASTER dmt: 144506

Date: 09/05/2018

OPERATOR NAME:
 KENNETH W SMITH
 Permit #: 35604
 Operator Agency/Dept: WARREN CO. SO
 Reference Standard #: 18170

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:44
INTERNAL STANDARD	VERIFIED	17:44
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	17:44
BLANK TEST	0.00	17:45
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	17:46
BLANK TEST	0.00	17:47
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	17:47
BLANK TEST	0.00	17:48

PO
[Signature]
 9/5/18

