

SUPERVISOR TEST

WARREN COUNTY
SHERRIF'S OFFICE
DATAMASTER dmt: 144506

Date: 09/14/2018

OPERATOR NAME:
KENNETH W SMITH
Permit #: 35604
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 18170

Nominal = 0.10 g/210L

BLANK TEST	0.00	16:23
INTERNAL STANDARD	VERIFIED	16:23
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	16:23
BLANK TEST	0.00	16:24
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	16:25
BLANK TEST	0.00	16:25
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	16:26
BLANK TEST	0.00	16:27

Kent Smith #301

9/14/18

