

SUPERVISOR TEST


WARREN COUNTY
 SHERIFF'S OFFICE
 DATAMASTER dmt: 144506

Date: 11/08/2018

OPERATOR NAME:
 CHARLES B GEISLER
 Permit #: 33757
 Operator Agency/Dept: WARREN CO. SO
 Reference Standard #: 18230

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:18
INTERNAL STANDARD	VERIFIED	03:18
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	03:19
BLANK TEST	0.00	03:20
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	03:20
BLANK TEST	0.00	03:21
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	03:21
BLANK TEST	0.00	03:22



Charles B. Geisler