

SUPERVISOR TEST

WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 12/26/2018

OPERATOR NAME:
CHARLES B GEISLER
Permit #: 33757
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 18310

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:41
INTERNAL STANDARD	VERIFIED	03:41
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	03:42
BLANK TEST	0.00	03:43
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	03:43
BLANK TEST	0.00	03:44
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	03:44
BLANK TEST	0.00	03:45

