

SUPERVISOR TEST

WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 105208

Date: 02/04/2019

OPERATOR NAME:  
CHARLES B GEISLER  
Permit #: 33757  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 18360

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:33
INTERNAL STANDARD	VERIFIED	23:33
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	23:34
BLANK TEST	0.00	23:34
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	23:35
BLANK TEST	0.00	23:36
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	23:36
BLANK TEST	0.00	23:37

