

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 02/24/2019

OPERATOR NAME:
CHARLES B GEISLER
Permit #: 33757
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 18360

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:14
INTERNAL STANDARD	VERIFIED	23:14
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	23:15
BLANK TEST	0.00	23:16
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	23:16
BLANK TEST	0.00	23:17
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	23:17
BLANK TEST	0.00	23:18

