

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 03/09/2019

OPERATOR NAME:
CHARLES B GEISLER
Permit #: 33757
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 18360

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:03
INTERNAL STANDARD	VERIFIED	23:03
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	23:03
BLANK TEST	0.00	23:04
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	23:04
BLANK TEST	0.00	23:05
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	23:06
BLANK TEST	0.00	23:07

