

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 03/29/2019

OPERATOR NAME:
KENNETH W SMITH
Permit #: 35604
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 18360

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:17
INTERNAL STANDARD	VERIFIED	09:17
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:17
BLANK TEST	0.00	09:18
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	09:19
BLANK TEST	0.00	09:20
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	09:20
BLANK TEST	0.00	09:21

P.O. Kenneth W Smith #301

3/29/19

