

SOLUTION CHANGE

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 07/29/2019

OPERATOR NAME:
KENNETH W SMITH
Permit #: 35604
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 19130

Nominal = 0.10 g/210L

BLANK TEST	0.00	16:46
INTERNAL STANDARD	VERIFIED	16:46
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	16:46

Blank Error

20 [Signature] 4/301

7/29/19

