

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 10/23/2019

OPERATOR NAME:
CHARLES B GEISLER
Permit #: 33757
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 19180

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:22
INTERNAL STANDARD	VERIFIED	23:22
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	23:23
BLANK TEST	0.00	23:24
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	23:24
BLANK TEST	0.00	23:25
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	23:25
BLANK TEST	0.00	23:26

