

SOLUTION CHANGE

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 11/01/2019

OPERATOR NAME:
RYAN C SCHROECK
Permit #: 41316
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 19180

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:38
INTERNAL STANDARD	VERIFIED	15:38
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	15:38
BLANK TEST	0.00	15:39
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	15:39
BLANK TEST	0.00	15:40
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	15:41
BLANK TEST	0.00	15:41

RS/S

