

SUPERVISOR TEST

WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 105208

Date: 11/22/2019

OPERATOR NAME:  
KENNETH W SMITH  
Permit #: 35604  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 19180

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:58
INTERNAL STANDARD	VERIFIED	09:58
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	09:58
BLANK TEST	0.00	09:59
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:59
BLANK TEST	0.00	10:00
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	10:01
BLANK TEST	0.00	10:01

*P.O. #301*

*11/22/19*

