

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106606

Date: 03/29/2019

OPERATOR NAME:
KENNETH W SMITH
Permit #: 35604
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 18360

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:45
INTERNAL STANDARD	VERIFIED	10:45
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	10:45
BLANK TEST	0.00	10:46
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	10:47
BLANK TEST	0.00	10:48
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	10:48
BLANK TEST	0.00	10:49

P.O. [Signature] #301

3/29/19

