

SOLUTION CHANGE

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106606

Date: 04/02/2019

OPERATOR NAME:
RYAN C SCHROECK
Permit #: 41316
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 19030

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:59
INTERNAL STANDARD	VERIFIED	08:59
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	08:59
BLANK TEST	0.00	09:00
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	09:01
BLANK TEST	0.00	09:02
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	09:02
BLANK TEST	0.00	09:03

Handwritten signature/initials

