

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106606

Date: 04/10/2019

OPERATOR NAME:
CHARLES B GEISLER
Permit #: 33757
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 19030

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:03
INTERNAL STANDARD	VERIFIED	21:04
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	21:04
BLANK TEST	0.00	21:05
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	21:05
BLANK TEST	0.00	21:06
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	21:06
BLANK TEST	0.00	21:07

