

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106606

Date: 06/16/2019

OPERATOR NAME:
CHARLES B GEISLER
Permit #: 33757
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 19090

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:44
INTERNAL STANDARD	VERIFIED	20:44
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	20:44
BLANK TEST	0.00	20:45
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	20:46
BLANK TEST	0.00	20:47
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	20:47
BLANK TEST	0.00	20:48

