

SOLUTION CHANGE

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106606

Date: 07/17/2019

OPERATOR NAME:
RYAN C SCHROECK
Permit #: 41316
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 19090

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:09
INTERNAL STANDARD	VERIFIED	17:09
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	17:10
BLANK TEST	0.00	17:11
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	17:11
BLANK TEST	0.00	17:12
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	17:12
BLANK TEST	0.00	17:13

